

## Calibration Certificate

**Certificate Number:** H73-1011917

Page 1 (1)

**Customer:**

**Manufacturer:** Vaisala Oyj  
**Instrument:** VL-4000-405 Analog Logger  
**Serial Number:** 10104034

The analog input(s) of the above described logger was calibrated by comparing the logger readings to a reference Precision Source Measurement Unit.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
 Traceable to the SI units (International System of Units) via national metrology institutes.

**Measurement results after adjustment:**

Reference		Channel 1					Channel 2					Notes
Reference	Observed	Error	Uncertainty	Limit	Pass/Fail	Observed	Error	Uncertainty	Limit	Pass/Fail		
V	V	V	V	V	V	V	V	V	V	V		
0.500	0.500	0.000	0.002	± 0.0075	Pass	0.500	0.000	0.002	± 0.0075	Pass		
2.500	2.500	0.000	0.002	± 0.0075	Pass	2.500	0.000	0.002	± 0.0075	Pass		
5.000	5.000	0.000	0.002	± 0.0075	Pass	5.000	0.000	0.002	± 0.0075	Pass		


Reference		Channel 3					Channel 4					Notes
Reference	Observed	Error	Uncertainty	Limit	Pass/Fail	Observed	Error	Uncertainty	Limit	Pass/Fail		
V	V	V	V	V	V	V	V	V	V	V		
0.500	0.500	0.000	0.002	± 0.0075	Pass	0.500	0.000	0.002	± 0.0075	Pass		
2.500	2.500	0.000	0.002	± 0.0075	Pass	2.499	-0.001	0.002	± 0.0075	Pass		
5.000	5.000	0.000	0.002	± 0.0075	Pass	4.999	-0.001	0.002	± 0.0075	Pass		

<b>Reference(s):</b>	<b>Serial Number</b>	<b>Calibration Date</b>	<b>Certificate</b>	<b>Next Calibration</b>
NI PXIE-4141	LP 14476	2014-02-26	1250-307054397	2015-02-28

**Note(s):**  
 This is new instrument without before adjustment data

**Calibration Date:** November 24, 2014      **Next Calibration:** November 24, 2015

**Date:** November 24, 2014      **Ambient Condition(s):**

**Signature:**       22.8 °C ±0.1 °C  
 Digitally signed by VIKR  
 Date: 2014.11.24 13:36:56 +02:00  
 Reason: Calibration responsible  
 Location: Vaisala Oyj  
 Ville Kröger  
 Calibration technician      21.9 %RH ±0.2 %RH  
 1024.2 hPa ±0.4 hPa